

PharmLabs Dallas LLC Certificate of Analysis

2567 Valley View Ln, Dallas, TX 75234, United States | License: 2020001  
 ISO/IEC 17025:2017 Certification L20-89-5 | Accreditation #85368



Sample **2000 BSO Roll Cool 44115**

Sample ID TX210517-013 (2139)	Matrix Topical
Tested for Proleve Distribution LLC	
Sampled - Received May 17, 2021	Reported May 21, 2021
Analyses executed CAN	Unit Mass (g) 90.0

CAN - Cannabinoid Profile Analysis

Analyzed May 21, 2021 | Instrument HPLC-DAD | Method WI-32  
 Measurement Uncertainty at 95% confidence 10.0%

Analyte	LOD	LOQ	Result %	Result mg/g	Result mg/Unit
Cannabidivarinic acid (CBDVa)	2.0e-06	5.0e-06	ND	ND	ND
Cannabidivarin (CBDV)	3.0e-06	1.0e-05	ND	ND	ND
Cannabidiolic acid (CBDA)	3.0e-06	8.0e-06	ND	ND	ND
Cannabigerolic acid (CBGa)	3.0e-06	8.0e-06	ND	ND	ND
Cannabigerol (CBG)	5.0e-06	1.6e-05	0.33	3.35	301.14
Cannabidiol (CBD)	6.0e-06	1.7e-05	2.14	21.43	1928.61
Tetrahydrocannabivarin (THCV)	6.0e-06	1.7e-05	ND	ND	ND
Tetrahydrocannabivarinic acid (THCVa)	5.0e-06	1.5e-05	ND	ND	ND
Cannabinol (CBN)	3.0e-06	1.0e-05	NT	NT	ND
Cannabinolic acid (CBNa)	8.0e-06	2.6e-05	ND	ND	ND
Δ9-Tetrahydrocannabinol (Δ9-THC)	1.2e-05	3.6e-05	ND	ND	ND
Δ8-Tetrahydrocannabinol (Δ8-THC)	1.5e-05	4.5e-05	ND	ND	ND
Cannabicyclol (CBL)	1.3e-05	3.8e-05	NT	NT	ND
Δ9-Tetrahydrocannabinolic acid (THCa)	9.0e-06	2.8e-05	ND	ND	ND
Cannabichromene (CBC)	6.0e-06	1.9e-05	ND	ND	ND
Cannabichromenic acid (CBCa)	2.2e-05	6.7e-05	ND	ND	ND
Total THC (THCa * 0.877 + THC)			ND	ND	ND
Total CBD (CBDA * 0.877 + CBD)			2.14	21.43	1928.61
Total CBG (CBGa * 0.877 + CBG)			0.33	3.35	301.14
TOTAL CANNABINOIDS			2.48	24.78	2229.75

Sample photography



ND Not Detected  
 N/A Not Applicable  
 NT Not Reported  
 LOD Limit of Detection  
 LOQ Limit of Quantification  
 <LOQ Detected  
 >ULOL Above upper limit of linearity  
 CFU/g Colony Forming Units per 1 gram  
 TNTC Too Numerous to Count



Scan the QR code to verify authenticity.

Authorized Signature  
  
 Dr. Archana R. Parameswar,  
 Laboratory Director  
 Fri, 21 May 2021 10:42:55 -0500

